




DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

	MANUFACTURER:	CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
	MEDICAL DEVICE:	SPIROMETER , SP10
	CLASSIFICATION - ANNEX IX:	Class II a, Rule 10
	CONFORMITY ASSESSMENT ROUTE:	Annex II excluding chapter 4
<p>WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.</p>		
<p>STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.</p>		
	NOTIFIED BODY:	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY
	IDENTIFICATION NUMBER:	 0123
	(EC) CERTIFICATE(S):	<u>G1 050972 0050 Rev.02</u>
	EC REP	
	EUROPEAN REPRESENTATIVE:	Shanghai International Holding Corp. GmbH(Hamburg) Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING: 2011-01-14 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:	QINHUANGDAO, 2019-07-23
SIGNATURE:	 <hr style="width: 100%;"/> President

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	EN60601-1:1990+A1:1993+A2:1995	Medical electrical equipment- Part 1: General requirements for safety
2	EN 60601-1-2:2007	Medical electrical equipment- Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN60601-1-4:1996+A1:1999	Medical electrical equipment - Part 1-4: General requirements for safety - Collateral standard: Programmable electrical medical systems
4	EN 60601-1-6:2007	Medical electrical equipment-Part 1-6: General requirements for basic safety and essential performance-Collateral Standard: Usability
5	EN ISO23747:2009 (ISO 23747:2007)	Anaesthetic and respiratory equipment - Peak expiratory flow meters for the assessment of pulmonary function in spontaneously breathing humans
6	EN 62304:2006	Medical device software-Software life-cycle processes