


**DECLARATION OF CONFORMITY  
TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING  
MEDICAL DEVICES**

<b>MANUFACTURER:</b>	<b>CONTEC MEDICAL SYSTEMS CO., LTD</b> No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA
<b>MEDICAL DEVICE:</b>	Sleep apnea screen meter, RS01
<b>CLASSIFICATION - ANNEX IX:</b>	Class II a, Rule 10
<b>CONFORMITY ASSESSMENT ROUTE:</b>	Annex II excluding chapter 4
<p>WE, ( CONTEC MEDICAL SYSTEMS CO., LTD ) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.</p>	
STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.	
<b>NOTIFIED BODY:</b>	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 M NCHEN, GERMANY
<b>IDENTIFICATION NUMBER:</b>	 0123
<b>(EC) CERTIFICATE(S):</b>	<u>G1 050972 0050 Rev.03</u>
<b>EUROPEAN REPRESENTATIVE:</b>	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany

**START OF CE-MARKING:** 2014-02-11 (Date or Lot or serial number)

<b>PLACE, DATE OF DECLARATION:</b>	QINHUANGDAO, 2019-11-07
<b>SIGNATURE:</b>	 _____ President

# DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

## Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	IEC 60601-1:2012	Medical electrical equipment -- Part 1: General requirements for basic safety and essential performance
2	EN 60601-1-2:2007	Medical electrical equipment- Part 1-2: General requirements for safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN 60601-1-8:2007	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance -Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
4	ISO 80601-2-61: 2011	Medical electrical equipment —Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment
5	EN 60601-1-6:2010	Medical electrical equipment-Part 1-6:General requirements for basic safety and essential performance-Collateral Standard: Usability
6	EN 62366:2008	Medical devices - Application of usability engineering to medical devices
7	EN 60601-1-11:2010	Medical electrical equipment -- Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
8	EN 62304:2006	Medical device software-Software life-cycle processes